

Technical Data Sheet

Mandil FFP2/V

Order no.: 411 281

Product designation: Filtering Facepiece Mandil FFP2/V

Applied standard: DIN EN 149:2009 (DIN = German Institute for Standardization)

Utilization: Protection against firm and liquid particles of slightly toxic (resp. noxious) substances,

including exhalation valve.

Application limit: 12,5 times the NPF value.

The scope of protection is determined by the product information, technical standards and

valid application rules.

Description: The filtering facepiece Mandil consists of several material layers.

Materials: Several used materials are incinerated.

Particle filter: Microfiberglass/Melt Blown Harness: Textile elastic band

Nosepiece: Polyethylene (PE)/Bamboo strengthened (Metal)

Nose upholstery: Foam (inside)

Exhalation valve: Housing: Polypropylene; Disc: Rubber

Operating principle: Particles are filtered through the Melt Blown-Filter.

Weight: approx. 15 g

Particle separation efficiency: Test Aerosol: Sodium chloride: 94%

Minimum separation efficiency: Paraffin oil: 94%

Inhalation resistance: at 30 l/min, constant flow max. 0.7 mbar (according to EN 149)

at 95 l/min, constant flow max. 2.4 mbar (according to EN 149)

Exhalation resistance: at 160 l/min, constant flow max. 3.0 mbar (according to EN 149)

Marking: Trade mark of manufacturer (LOGO)

Product designation Applied standard

Expiry of shelf-life (MM/YYYY) or (YYYY) Consider the information brochure

Lot-No.

CE-Identification Notified body

Identification color: white

Handling: Take the Filtering Facepiece out of the packing directly before use.

Don the Filtering Facepiece over the mouth, nose and chin in accordance with the instructions

(See retail packaging)

Storage conditions: Ambient temperature. Protect against cold, heat and humidity.

Consider conditions noted on the packaging.

Tel.:

Fax:

Warning reference: Please consider the information brochure!

User references: EKASTU Safety GmbH guarantees the indicated achievement according to class and type.

It is to be noted that the laboratory test values can considerably deviate from those, which are reached in practice. The user must read and understand all functional information. Use the Respiratory protective devices with this information brochure, the relevant valid statutory regulations and the safety requirements of the profession associations, particularly the regulations for use in accordance with DGUV Rule 112-190 resp. DIN EN 529 'Respiratory protective devices – recommendation for selection, use, care and maintenance – guidelines'.

+49/7151/975099-0

+49/7151/975099-30

TD 411281 English 07-07.17

info@ekastu.de

www.ekastu.de